



CANADIAN DECLARATION OF CONFORMITY

Villari B.V.
Oostsingel 209
2612 HL, Delft
The Netherlands

This declaration of conformity is issued for:

Product Name:	Villari Ferro One rigid sensor probe
Trademark:	Villari Ferro One 75mm rigid sensor probe
Model:	IN-SP-075R
Serial number of product covered by this CAN DoC	IN-SP-075R-DD-EE-FF-XXXXXXX - DD = Major revision - EE = Region (e.g. 00 = global) - FF = Production batch - XXXXXXX = individual identifier
Accessories bundled with the product	Villari Ferro One system

We, Villari B.V. declare under our sole responsibility that the above named product(s) is in conformity with the applicable Canadian regulations and standards

Health & Safety (Electrical safety)	CAN/CSA C22.2 No. 61010-1-12; UPD1:2015; UPD2:2016; AMD1:2018; COR1:2019; UPD3:2023; UPD4:2024
Electromagnetic Compatibility (EMC)	ICES-003 Issue 7
IP Rating – IP67	IEC-EN 60529:1991+A1:2000+A2:2013

The accredited testing facility, DEKRA Certification B.V., located at Meander 1051, 6825 MJ, Arnhem, performed international and additional Canadian type of examination issuing following certificate number(s) and test report(s):

- **NL-119312 (Health and safety)**
- **Test report no. 2297499.0501-EMC (EMC)**
- **Test report no. 2297638.50E (IP67)**

Signed for and on behalf of Villari B.V.
Date of issue: **16 sep. 2025**

Olivier Baas

Olivier Baas, CEO

Floris Achterberg

Floris Achterberg, COO/CFO