



EU DECLARATION OF CONFORMITY

Villari B.V.
Oostsingel 209
2612 HL, Delft
The Netherlands

This declaration of conformity is issued for:

Product Name:	Villari Ferro One rigid sensor probe
Trademark:	Villari Ferro One 75mm rigid sensor probe
Model:	IN-SP-075R
Serial number of product covered by this EU DoC	IN-SP-075R-DD-EE-FF-XXXXXXX - DD = Major revision - EE = Region (e.g. 00 = global) - FF = Production batch - XXXXXXX = individual identifier
Accessories bundled with the product	Villari Ferro One system

We, Villari B.V. declare under our sole responsibility that the above named product(s) conform to the essential requirements of the following European Union directives:

The object of the declaration described above is in conformity with the following regulations:

Health & Safety (Low Voltage Directive)	2014/35/EU
Electromagnetic Compatibility (EMC) Directive	2014/30/EU
RoHS 2 Directive	2011/65/EU - 2015/863/EU
WEEE Directive	2012/19/EU

The following harmonised standards and normative documents are those to which the product's conformance is declared, and by specific reference to the essential requirements of the referenced Directives:

Health & Safety	IEC-EN 61010-1:2010
EMC	IEC-EN 61000-6-2:2019 IEC-EN 61000-6-4:2019
IP Rating – IP67	IEC-EN 60529:1991+A1:2000+A2:2013

The accredited testing facility, DEKRA Certification B.V., located at Meander 1051, 6825 MJ, Arnhem, conducted the conformity assessments and issued the following certificates: **NL-118884 (EMC)**, **NL-119312 (Health and Safety)**, and test report no. **2297638.50E (IP67 rating)**.

Signed for and on behalf of Villari B.V.

Date of issue: **16 sep. 2025**

Olivier Baas

Olivier Baas, CEO

Floris Achterberg

Floris Achterberg, COO/CFO